



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that PPD Global Central Lab, using Tosoh G8 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>G8</b>	Calibrator Lot: <b>ZS7001, ZS7001</b>	Column Lot: <b>N</b>
Reagent Lot: <b>N154D, N8253D, N8-353D, HW-530-LU</b>	Calibrator Assigned Values: <b>5.7%, 10.9%</b>	

Date of Certification: May 1, 2019

Certification Expires: May 1, 2020

*Daniel Soule* NGSP Steering Committee Chair     
 *Rondie R Little PhD* NGSP Network Coordinator     
 *Shawn Smith* SRL director/ supervisor